

Declaration of conformity

*Name and adress
of Manufacturer:* Roboscreen GmbH
Hohmannstrasse 7
04129 Leipzig,
Deutschland

SRN of Manufacturer: DE-MF-000024661

The **Roboscreen GmbH** declares on own responsibility that the medical device

Name: INSTANT Virus RNA/DNA Kit

Catalogue number: 847-0259200602, 847-0259200603

Description: Extraction of viral nucleic acids by means of DC-technology

Basic-UDI-DI: 426238673025920HW

Risk class: Class A

meets all applicable requirements of the Regulation (EU) 2017/746 on in vitro diagnostic medical devices. In the context of conformity with the mentioned regulation the standards referred to in Chapter 4.1 of the Technical Documentation have been fully or partially applied.

*Applied
conformity assessment
procedure:* Annex II (Technical Documentation) and Annex III (Post-market
surveillance)

Applied CS: None

*Name and referenc number of
notified body:* Not applicable

*Certificate issued by the
notified body:* Not applicable

Validity (DD.MM.YYYY): 02.08.2022 – 25.05.2023

Place: Leipzig

Date: 02.08.2022

Name and Function: Dr. Ingolf Lachmann, CEO

Legally binding signature:



Roboscreen GmbH
Hohmannstraße 7
04129 Leipzig • Germany

Phone: +49 341 989 734 0
Fax: +49 341 989 734 199
info@roboscreen.com
www.roboscreen.com

CEO
Dr. Ingolf Lachmann

Ust-IdNr. DE 813064314
County Court Leipzig
HRB 17503
Leipzig